



**Clinical Study for  
SARS-CoV-2 & Flu AB & RSV &  
Adenovirus Antigen Combo Rapid Test**

**By: Printed Name**

Fangli Tong

**Signature**

Fangli Tong

**Date**

2022.01.29

## I. Instruction

The SARS-CoV-2 & Flu A/B & RSV & Adenovirus Antigen Combo Rapid Test is a rapid test for the qualitative and differential detection of SARS-CoV-2, Influenza virus type A and type B nucleoprotein antigens, respiratory syncytial virus(RSV) and respiratory adenovirus antigens in nasal swab or nasopharyngeal swab.

## II. Study Objective

To perform a clinical sensitivity, specificity study.

## III. Clinical Study Site and Study Period

Sample collection sites in China	Testing sites in China
<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China	<u>Site 1:</u> <u>Shenzhen CDC</u> No. 8 Longyuan Road, Nanshan District, Shenzhen, P.R. China
<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China	<u>Site 2:</u> Adicon No.208 Zhenzhong Road, West Lake District, Hangzhou, Zhejiang, P.R. China

## IV. Study Procedure

### Material :

Nasal/Nasopharyngeal swab samples from infected patients and non-infected patients.

SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test: Lot#: CFRAC1060002 Extraction Solution, Lot#: 210520022

Comparison Test:

- (1) RT-PCR, Novel Coronavirus (2019-nCoV) Nucleia Acid Diagnostic Kit (PCR-Fluorescence Probing), CE Marked, manufactured by Sansure BioTech Inc. Promotor SARS-COV-2 RT-PCR Test Kit, CE Marked, manufactured by ACON Biotech(Hangzhou)Co., Ltd.
- (2) SD BIOLINE Influenza Antigen, CE Marked product, manufactured by Standard Diagnostics, Inc.
- (3) Genesis Respiratory syncytial virus Antigen Rapid Diagnostic Test Kit, CE Marked product, manufactured by Hangzhou Genesis Biodetection & Biocontrol Co., Ltd.
- (4) Genesis Adenovirus Antigen Rapid Test Kit, CE Marked product, manufactured by Hangzhou Genesis Biodetection & Biocontrol Co., Ltd.

**Method :**

- 1) Collected at least 100 SARS-CoV-2 positive nasal/nasopharyngeal specimens, at least 30 Influenza A positive nasal/nasopharyngeal specimens, at least 30 Influenza B positive nasal/nasopharyngeal specimens, at least 30 RSV nasal/nasopharyngeal specimens, at least 30 adenovirus nasal/nasopharyngeal specimens. All specimens are from symptomatic individuals who were SARS-CoV-2 or Influenza A or Influenza B or RSV or adenovirus positive with comparison test.
- 2) Collected at least 100 negative clinical specimens from symptomatic individuals.
- 3) Test the specimens with SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test follow the package insert.

**V. Acceptance Criteria :**

- 1) For SARS-CoV-2  
Sensitivity:  $\geq 85\%$ ; Specificity:  $\geq 95\%$
- 2) For Influenza A/B  
Sensitivity:  $\geq 80\%$ ; Specificity:  $\geq 95\%$
- 3) For RSV  
Sensitivity:  $\geq 85\%$ ; Specificity:  $\geq 95\%$
- 4) For Adenovirus  
Sensitivity:  $\geq 85\%$ ; Specificity:  $\geq 95\%$

**VI. Test Result :****1. Nasal Swab Sample:****(1) For SARS-CoV-2**

Candidate method		RT-PCR		
		Negative	Positive	Total
SARS-CoV-2 Test Results	Negative	403	4	407
	Positive	2	99	101
	Total	405	103	508

Relative Sensitivity: 96.12% (95% CI: 90.12%-98.80%)

Relative Specificity: 99.51% (95% CI: 98.10%-99.99%)

Accuracy: 98.82% (95% CI: 97.38%-99.52%)

**(2) For Influenza A**

Candidate method		Comparato		
		Negative	Positive	Total
Flu A Test Results	Negative	461	0	461
	Positive	2	45	47
	Total	463	45	508

Relative Sensitivity: 100.00% (95% CI: 90.62%-100.00%)

Relative Specificity: 99.57% (95% CI: 98.33%-99.99%)

Accuracy: 99.61% (95% CI: 98.48%-99.99%)

(3) For Influenza B

Candidate method		Comparato		
		Negative	Positive	Total
Flu B Test Results	Negative	457	1	458
	Positive	2	48	50
	Total	459	49	508

Relative Sensitivity: 97.96% (95% CI: 88.31%-99.99%)

Relative Specificity: 99.56% (95% CI: 98.32%-99.99%)

Accuracy: 99.41% (95% CI: 98.19%-99.88%)

(4) For Respiratory Syncytial Virus

Candidate method		Comparato		
		Negative	Positive	Total
RSV Test Results	Negative	470	0	470
	Positive	3	35	38
	Total	473	35	508

Relative Sensitivity: 100.00% (95% CI: 88.24%-100.00%)

Relative Specificity: 99.37% (95% CI: 98.06%-99.88%)

Accuracy: 99.41% (95% CI: 98.19%-99.88%)

(5) For Adenovirus

Candidate method		Comparato		
		Negative	Positive	Total
Adenovirus Test Results	Negative	465	1	466
	Positive	3	39	42
	Total	468	40	508

Relative Sensitivity: 97.50% (95% CI: 85.96%-99.99%)

Relative Specificity: 99.36% (95% CI: 98.04%-99.87%)

Accuracy: 99.21% (95% CI: 97.92%-99.77%)

**2. Nasopharyngeal Swab Sample:**

(1) For SARS-CoV-2

Candidate method		RT-PCR		
		Negative	Positive	Total
SARS-COV-2 Test Results	Negative	401	4	405
	Positive	2	97	99
	Total	403	101	504

Relative Sensitivity: 96.04% (95% CI: 89.93%-98.77%)

Relative Specificity: 99.50% (95% CI: 98.09%-99.99%)

Accuracy: 98.81% (95% CI: 97.36%-99.52%)

(2) For Influenza A

Candidate method		Comparato		
		Negative	Positive	Total
Flu A Test Results	Negative	471	0	471
	Positive	3	30	33
	Total	474	30	504

Relative Sensitivity: 100.00% (95% CI: 86.53%-100.00%)

Relative Specificity: 99.37% (95% CI: 98.07%-99.88%)

Accuracy: 99.40% (95% CI: 98.18%-99.88%)

(3) For Influenza B

Candidate method		Comparato		
		Negative	Positive	Total
Flu B Test Results	Negative	471	1	472
	Positive	2	30	32
	Total	473	31	504

Relative Sensitivity: 96.77% (95% CI: 82.42%-99.99%)

Relative Specificity: 99.58% (95% CI: 98.37%-99.99%)

Accuracy: 99.40% (95% CI: 98.18%-99.88%)

(4) For Respiratory Syncytial Virus

Candidate method		Comparato		
		Negative	Positive	Total
RSV Test Results	Negative	471	0	471
	Positive	3	30	33
	Total	474	30	504

Relative Sensitivity: 100.00% (95% CI: 86.53%-100.00%)

Relative Specificity: 99.37% (95% CI: 98.07%-99.88%)

Accuracy: 99.40% (95% CI: 98.18%-99.88%)

(5) For Adenovirus

Candidate method		Comparator		
		Negative	Positive	Total
Adenovirus Test Results	Negative	469	1	470
	Positive	3	31	34
	Total	472	32	504

Relative Sensitivity: 96.88% (95% CI: 82.89%-99.99%)

Relative Specificity: 99.36% (95% CI: 98.06%-99.87%)

Accuracy: 99.21% (95% CI: 97.90%-99.77%)

## **VII. Conclusion:**

Nasal specimens: for SARS-CoV-2, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 96.12%, specificity of 99.51%, and accuracy of 98.82% when comparing with RT-PCR. For Influenza A, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 100.00%, specificity of 99.57%, and accuracy of 99.61% when comparing with Rapid Test Method. For Influenza B, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 97.96%, specificity of 99.56%, and accuracy of 99.41 % when comparing with Rapid Test Method. For Respiratory Syncytial Virus, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 100.00%, specificity of 99.37%, and accuracy of 99.41 % when comparing with Rapid Test Method. For Adenovirus, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 97.50%, specificity of 99.36%, and accuracy of 99.21% when comparing with Rapid Test Method.

Nasopharyngeal specimens: for SARS-CoV-2, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 96.04%, specificity of 99.50%, and accuracy of 98.81% when comparing with RT-PCR. For Influenza A, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 100.00%, specificity of 99.37%, and accuracy of 99.40% when comparing with Rapid Test Method. For Influenza B, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 96.77%, specificity of 99.58%, and accuracy of 99.40% when comparing with Rapid Test Method. For Respiratory Syncytial Virus, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 100.00%, specificity of 99.37%, and accuracy of 99.40 % when comparing with Rapid Test Method. For Adenovirus, the SARS-CoV-2 & Flu AB & RSV & Adenovirus Antigen Combo Rapid Test has sensitivity of 96.88%, specificity of 99.36%, and accuracy of 99.21% when comparing with Rapid Test Method.